



For veterinary use only

INTENDED USE

The Vcheck Canine Progesterone is an *in vitro* diagnostic test kit for the quantitative measurement of progesterone concentration in canine serum. Progesterone is a hormone produced by the ovaries and placenta that helps prepare and maintain the uterus for pregnancy. In non-pregnant dogs, the concentration of progesterone is low in their serum. Progesterone levels begin to rise around the time of LH (Luteinizing Hormone) surge. Thus, the production of progesterone starts prior to ovulation at the end of proestrus and reaches increased levels at the time of ovulation. The measurement of Progesterone can be used to determine optimal breeding dates, predict parturition dates or time a Cesarean section and detect reproductive disorders. This is designed to be used only by veterinarians.

PRINCIPLE

The Vcheck Canine Progesterone Test Kit is based on competitive immunoassay method for the quantitative measurement of canine progesterone concentration. When the specimen is delivered to the sample hole of the test device, progesterone in the specimen and the colloidal gold-labeled anti-progesterone monoclonal antibody in conjugate pad migrate along the nitrocellulose membrane. They react with progesterone-BSA coated on the membrane. Here, the non-binding gold-labeled antibody binds to progesterone-BSA on the membrane. The density of the test line is inversely proportional to the canine progesterone concentration in canine serum. The BIONOTE Vcheck Analyzer reads the density of the test line and calculates the progesterone concentration from the calibration curve data. The control line is a reference line that indicates the test has been performed correctly.

MATERIALS PROVIDED

Reagent	10 Tests/Kit
① Vcheck cProgesterone Test device	10
② Assay diluent tube	10
③ Disposable pipette tip	20
④ Instructions for use	1

MATERIALS REQUIRED, BUT NOT PROVIDED

- BIONOTE Vcheck Analyzer
- 50 µl pipette
- 100 µl pipette

STORAGE AND STABILITY

- Store the test kit at 2~8°C. **DO NOT FREEZE.**
- Do not store the test kit in the direct sunlight.
- The test kit is stable until the expiry date that is marked on the package label.

PRECAUTIONS

- The test kit is for canine use only. Do not use for other animals.
- The test device is sensitive to humidity and heat. Perform the test immediately after removing the test device from the foil pouch.
- Do not reuse test components.
- Do not touch the membrane in the result window of the test device.
- Do not use the test kit beyond the stated expiry date marked on the label.
- Do not use the test kit if the pouch is damaged or the seal is broken.
- Do not mix components from different lot numbers, the components in this kit have been quality control tested as a standard batch unit.
- All samples should be handled as being potentially infectious. Wear protective gloves while handling samples. Wash hands thoroughly afterwards.

- Decontaminate and dispose of all samples, used kits and potentially contaminated materials safely in accordance with national and local regulations.
- Hemolyzed samples or serum with precipitate such as fibrin may give erroneous results.

COLLECTION AND PREPARATION OF SAMPLE

- Canine serum should be used with this test. A method for preparing the serum is as follows.
- [Serum]** Collect the whole blood into a blood collection tube containing NO anticoagulant. Leave to settle for 30 minutes for blood coagulation and then centrifuge to obtain the serum supernatant.
- If serum samples are not tested immediately, they should be refrigerated at 2~8°C and used within 7 days. For longer storage, serum can be frozen (-20°C or colder). Frozen samples should be brought to room temperature (15~30°C) prior to use.

TEST PROCEDURE

All reagents and samples must be at room temperature (15~30°C) before use.

[Coding]

- Turn on V200 Analyzer and select "Standard Test".



- Remove the test device from the foil pouch. Once the "insert device" is displayed in the screen, insert the test device.




[Dilution of sample & Measurement]

- Use a 50 µl pipette to draw 50 µl of the sample and add to the assay diluent tube (100 µl).



- Use a 50 µl pipette to mix the sample with diluent by pipetting for 5~6 times.



- Add 100 µl of mixture in the sample hole of test device using 100 µl pipette and press START.



- The V200 analyzer will display the test result on the screen after 15 minutes.
- Remove the test device.



INTERPRETATION OF THE RESULT

- Read the concentration value of canine Progesterone appearing on the display of the BIONOTE Vcheck Analyzer. (1~30 ng/ml)
- If "<1 ng/ml" appears on the display, it means the concentration of canine progesterone the specimen is less than 1 ng/ml.
- If ">30 ng/ml" appears on the display, it means the concentration of canine progesterone the specimen is greater than 30 ng/ml.
- 1 ng/ml is equal to 3.18 nmol/L

REFERENCE RANGE

< 1.0 ng/ml	Anestrus or proestrus
1.0-1.99 ng/ml	Pre-LH surge
2.0-2.99 ng/ml	LH surge
3.0-4.99 ng/ml	Post-LH surge, Pre-ovulation
5.0-12.0 ng/ml	Ovulation (It may vary with breed and size)
> 12.0 ng/ml	Diestrus

SCREEN MESSAGES AND TROUBLE SHOOTING

[V200]

Error message	Error description
Contaminated Device	The test device is damaged or inserted improperly. Solution: Discard the test device and retest with a new test device and a new specimen.
Insufficient Sample	An insufficient amount of blood has been applied. Solution: Retest with a new test device with enough specimen, ensuring that blood is placed in to the narrow channel in the top edge of the test device.
Expired Device	The test devices are expired. Solution: Retest with a new test device that is not expired.
Temperature Error	The environmental temperature is above or below the operating range of the analyzer. Solution: Move to an area in the acceptable temperature range for the analyzer and perform the test. Do not heat or cool the analyzer artificially.
Printer Connection Fail	The communication between analyzer and barcode or printer has failed. Solution: Reconnect the analyzer and external device. If the error continues after turning ON/OFF the analyzer, please contact BioNote, Inc.
Barcode Error	The measured total hemoglobin is out of the range of 7 to 23 g/dL. Solution: This error occurs when a specimen has a total hemoglobin in the abnormal range. If the error continues after turning ON/OFF the analyzer, please contact BioNote, Inc.
Extremely Total hemoglobin	The test is invalid. Solution: Retest with a new test device and a new patient specimen. If the error continues after turning ON/OFF the analyzer, please contact BioNote, Inc.
Result: Invalid	The calibration is overdue. Solution: If the error continues after turning ON/OFF the analyzer, please contact BioNote, Inc.
Calibration Overdue	Loading a test device that is not supported by the analyzer. Solution: Check whether the test device is manufactured by BioNote, Inc.
Not Supported Device	Internal error has occurred. Solution: If the error continues after turning ON/OFF the analyzer, please contact BioNote, Inc.
EEE	

제품명

내분비물질검사시약 [2]

형 명

Vcheck cProgesterone

제품의 구성품 및 포장 단위

구성품	5 Tests/Kit	10 Tests/Kit
검사용 디바이스	5	10
검체희석액	5	10
파이펫 팁	10	20
사용 설명서	1	1

사용 목적

Vcheck cProgesterone은 개의 혈청에서 Canine Progesterone을 면역 측정법으로 정량하는 동물용 체외진단 분석기용 시약으로 동물병원 및 병성감정 기관에서만 사용 가능합니다. (일반인 사용 불가)

사용 방법

[검체준비 및 저장방법]

- 개의 혈청을 검체로 사용하며, 검체 준비방법은 다음과 같습니다.
- 1) 멸균된 주사기로 채혈한 혈액을 항응고제가 들어 있지 않은 튜브에 수집합니다.
- 2) 실온에 약 30분간 방치해 응고가 일어나도록 한 후 원심분리에 의해 상층의 혈청을 분리시킵니다.
- 3) 분리된 혈청은 2~8°C에서 보관시 1주간 사용가능하며, 장기 보관이 필요한 경우 -20°C 이하에서 냉동 보관합니다.
- 4) 혈구가 깨져 완전히 분리되지 않은 혈청은 검사결과에 영향을 줄 수 있으므로 검체준비에 유의하도록 합니다.

* 2~8°C 혹은 -20°C이하에서 보관된 검체는 반드시 상온에 노출시킨 후 사용합니다.

[검사 전 유의사항]


- 1) 검사용 디바이스는 온도 및 습기에 민감하므로 호일 파우치에 포장된 상태로 실온에 30분 적응시킨 후 시험 시작 직전에 호일 파우치에서 꺼냅니다.
- 2) 검사용 디바이스의 검사창은 손으로 만지지 않습니다.
- 3) 검사용 디바이스가 포함되어있는 호일 파우치가 손상되었을 경우 사용하지 않습니다.
- 4) 키트의 구성물을 다른 로트와 혼용하여 사용하지 않습니다.

[검사과정]

모든 키트 구성품과 검체는 시험 전에 반드시 상온(15~30°C)에 최소 30분 정도 노출시킨 후 사용합니다.

디바이스 확인

- 1) BIONOTE 브이체크 분석기의 전원을 키고 Standard Mode를 누릅니다.



- 2) 호일 파우치를 개봉하여 검사용 디바이스를 꺼내고, 기기에 삽입하여 기기가 디바이스의 바코드를 먼저 읽을 수 있도록 합니다.



검체 희석 및 측정

- 1) 50 µl 피펫을 이용하여 분리한 혈청 50 µl를 취합니다.



- 2) 채취한 검체를 검체희석액(100 µl)이 들어있는 희석액 튜브에 넣은 후 충분히 섞일 수 있도록 50 µl 피펫으로 5~6번 섞어줍니다.



- 3) 100 µl 피펫을 이용하여 샘플혼합액 100 µl를 검체 점적 부위에 점적합니다.
- 4) 기기의 'START' 버튼(전원 버튼)을 눌러 검사를 시작합니다.



- 5) 15분 후, 기기 표시창에 표시된 결과를 확인합니다.



- 6) 측정이 완료되면, 사용했던 검사용 디바이스를 기기에서 제거합니다.

*검체량 및 반응시간을 다르게 할 경우 시험결과에 영향을 줄 수 있으므로 정해진 검체량(50 µl)과 시간(15분)을 지켜 시험하도록 합니다.

[결과판정]

- 1) Canine Progesterone 농도 확인: 기기 결과 창에 농도가 표시된다.(1~30 ng/ml)
- 2) <1 ng/ml이 나온다면 검체농도가 1 ng/ml 미만입니다.
- 3) >30 ng/ml이 나온다면 검체농도가 30 ng/ml 초과입니다.
- 4) 1 ng/ml는 3.18 nmol/L입니다.

사용 시 주의사항

- 1) 개의 체외진단용으로만 사용합니다.
- 2) 본 시약은 2~8°C에서 보관이 필요한 제품으로, 사용 시 키트에 포함된 각종시약을 실온에 충분히 적응시킨 후 검사를 진행 하도록 합니다.
- 3) 보관 중인 디바이스가 습기에 노출되면 제품의 성능이 저하될 수 있으므로 사용 직전에 개봉하고, 개봉 후 10분 이내에 사용합니다.
- 4) 반드시 각 검체마다 별개의 1회용 구성품을 사용합니다.
- 5) 검사 디바이스의 표시창 내 멤브레인을 손으로 만지는 등의 직접적인 접촉은 검사 결과에 영향을 미칠 수 있습니다.
- 6) 검체에 혈구가 남아있거나 용혈이 일어난 검체는 부정확한 결과가 나타날 수 있으므로 깨끗이 분리된 혈청만 사용하도록 합니다.
- 7) 검체량(50 µl) 및 반응시간(15분)을 해당 사용방법대로 지키지 않을 경우 부정확한 결과가 나올 수 있으므로 검체량 및 반응시간을 준수하여 시험 하도록 합니다.
- 8) 사용 기간이 경과한 시약은 사용하지 않습니다.
- 9) 검체는 미지의 바이러스나 세균 감염원으로서의 위험성을 내포하고 있으므로 취급에 주의하며, 감염 가능한 물질의 취급 시에는 일회용 장갑을 사용하고 취급 후 손을 깨끗이 씻습니다.
- 10) 실험에 사용한 고형 폐기물은 121°C에서 15분 이상 고압 증기 멸균하여 폐기합니다.
- 11) 본 제품은 개 Progesterone에 관한 간이 정량 목적으로 고안된 시약으로 간편하고 신속한 방법으로 결과를 얻을 수 있으나, 보다 정밀한 원리로 고안된 검사법과 검출 감도 면에서 차이를 보일 수 있습니다.
- 12) 전문 수의사가 본 제품의 결과 및 다른 검사 결과와 임상 소견에 근거하여 최종 진단을 내려야 합니다.
- 13) BIONOTE 브이체크 분석기는 15~30°C에서 사용을 권장합니다.

저장방법 및 사용기한

구성 시약	개봉 여부	보관 조건	사용 기간	비 고
검사용 디바이스	미개봉 개봉	온도 2~8°C, 밀봉 보관 안됨	12 개월 -	완제품 즉시사용
검체희석액	미개봉 개봉	온도 2~8°C, 밀봉 보관 안됨	12 개월 -	완제품 즉시사용

※ 검사키트는 냉동보관하지 않도록 주의.

*참고) 오류 메시지, 원인 및 조치 사항 [V100]

오류 메시지	원인 및 조치 사항
INVALID DEVICE	테스트가 유효하지 않음 조치 사항 : 새로운 디바이스 혹은 새로운 검체를 이용하여 재검사 및 기기 다시 시작
CONTAMINATED	손상된 디바이스 사용 및 부적절한 디바이스 삽입 조치 사항 : 새로운 디바이스를 이용하여 다시 검사
SAMPLE NOT DETECTED	검체 점적량 부족 조치 사항 : 새로운 디바이스에 알맞은 양의 검체를 적용
EXPIRED	디바이스 유효기한이 지남 조치 사항 : 유효기한이 지나지 않은 새 디바이스를 이용하여 다시 검사
TEMPERATURE	기기 혹은 디바이스의 온도가 매우 낮거나 높은 경우 조치 사항 : 기기 혹은 디바이스를 상온에 적정시간 인큐베이션한 후 검사
BARCODE	기기가 디바이스 바코드를 읽지 못함 조치 사항 : 기기 다시 시작
DEVICE	기기가 디바이스를 감지하지 못함 조치 사항 : 디바이스 다시 삽입 및 기기 다시 시작
PC	기기가 PC를 감지하지 못함 조치 사항 : PC 다시 연결 및 기기 다시 시작
PRINTER	기기가 외부 프린터를 감지하지 못함 조치 사항 : 외부 프린터 다시 연결 및 기기 다시 시작
EEE	내부 시스템 문제 조치 사항 : 기기 다시 시작

[V200]

오류 메시지	원인 및 조치 사항
Contaminated Device	손상된 디바이스 사용 및 부적절한 디바이스 삽입 조치 사항 : 새로운 디바이스 혹은 새로운 검체를 이용하여 다시 검사
Insufficient Sample	검체 점적량 부족 조치 사항 : 충분한 양의 검체로 새 디바이스를 이용하여 다시 검사
Expired Device	디바이스 유효기한 지남 조치 사항 : 유효기한이 지나지 않은 새 디바이스를 이용하여 다시 검사
Temperature Error	기기 혹은 디바이스의 온도가 매우 낮거나 높은 경우 조치 사항 : 기기 혹은 디바이스를 상온에 적정시간 인큐베이션한 후 검사
Printer Connection Fail	기기가 외부 프린터를 감지하지 못함 조치 사항 : 외부 프린터 재연결 및 기기 다시 시작
Barcode Error	기기가 바코드를 읽지 못함 조치 사항 : 기기 다시 시작
Extremely Total hemoglobin	측정된 총 헤모글로빈의 수치가 7 g/dL 이상 23 g/dL 이하 범위를 벗어나는 경우 조치 사항 : 기기 다시 시작
Result: Invalid	테스트가 유효하지 않음 조치 사항 : 새로운 디바이스 혹은 새로운 검체를 이용하여 재검사 및 기기 다시 시작
Calibration Overdue	검교정 기한이 지남 조치 사항 : 기기 다시 시작
Not Supported Device	기기가 지원하지 않는 디바이스 조치 사항 : BIONOTE에서 제조된 디바이스인지 확인
EEE	내부 시스템 문제 조치 사항 : 기기 다시 시작